ZC0016

ISO calibration certificate for temperature sensors 3-point



ISO-Kalibrierung Temp. 3-Pkt Werkskalibrierschein ifm electronic Factory calibration certificate Messumformer-Typ: 1000 Auftragsnummer: -Referenzgerät: Serien-Nr.: reference unit: Messbereich: 3.80075 Prüfnormal, Temperatur: temperature range testing device, temperature: 1.000 Serien-Nr.: Ausgangssignal: serial no.: Genauigkeit: 17000 000 PTB-Zertifikat: PTB certificate Kalibrierdatum: 1000 Umgebungstemperatur: date of calibration: the person grade

Mechanical data			
Weight	[g]	19.171	
Remarks			
Remarks		the ISO calibration certificate is not recognised according to the EA MLA (Multilateral Agreement) and is therefore considered as not traced.	
		The calibration certificate must be ordered in conjunction with a sensor.	
		The corresponding sensor is delivered along with the calibration certificate.	
		The calibration certificate is in two languages (German/English).	
		Subsequent orders are only possible if the sensor is in a clean condition upon return.	
		In the internet version of the data sheet a sample certificate is stored in the tab "Downloads".	

ZC0016

ISO calibration certificate for temperature sensors 3-point



ISO-Kalibrierung Temp. 3-Pkt

Other data			
Measurement points (to ISO 9110)			
1	65 °C		
2	85 °C		
3	123 °C		