

## ISO KALIBRIERUNG SM ATM GERÄTE

Application	
Design	for type SM6x2x, SM7x2x, SM8x2x, SM8x3x
Mechanical data	
Weight	[g] 33.17
Remarks	
Remarks	The calibration certificate must be ordered in conjunction with a sensor.
	The corresponding sensor is delivered along with the calibration certificate.
	The calibration certificate is in two languages (German/English).
	Recalibration or subsequent orders are only possible if the device is returned in a clean condition, without any adaptations attached (condition on delivery) and accompanied by a completed Declaration of Decontamination.
	In the internet version of the data sheet a sample certificate is stored in the tab "Downloads"